Accurate frequency measurement of a submillimeter wave gyrotron output using a far-infrared laser as a reference

著者	IDEHARA T, MITSUDO S, SAITO S, OGAWA I, OKAJIMA S
journal or publication title	Review of Scientific Instruments
volume	74
number	5
page range	2860-2862
year	2003-05
URL	http://hdl.handle.net/10098/1587

doi: 10.1063/1.1568549

Accurate frequency measurement of a submillimeter wave gyrotron output using a far-infrared laser as a reference

T. Idehara, S. Mitsudo, and S. Saito

Research Center for Development of Far-Infrared Region, Fukui University, Bunkyo 3-9-1, Fukui 910-8507, Japan

I. Ogawa

Cryogenic Laboratory, Faculty of Engineering, Fukui University, Bunkyo 3-9-1, Fukui 910-8507, Japan

S. Okajima

College of Engineering, Chubu University, Kasugai, Aichi 487-8501, Japan

Accurate measurement of gyrotron frequency has been carried out by using a far-infrared (FIR) molecular laser as a reference. Highly stable operation of a FIR laser enables us to make the resolution of frequency measurement so precisely that both the frequency width Δf and the frequency fluctuation level δf can be studied within the accuracy of 1 kHz or less. When our submillimeter wave gyrotron operates in cw mode, the observed frequency width Δf is around 2 kHz and the fluctuation level δf 10 kHz. Such an accurate measurement is a useful tool for development of high quality gyrotrons with stable frequency.

I. INTRODUCTION

Gyrotron FU series developed in Research Center for Development of Far-Infrared Region, Fukui University (FIR FU) is one of high frequency, step-tunable gyrotrons. The gyrotron series has achieved high frequency operations up to 889 GHz (the corresponding wavelength is 337 μm),¹ which is the current world record. The series is being applied to some fields, for example, plasma scattering measurement of a drift wave excited in a plasma of a helical system,² electron spin resonance spectroscopy for material science,³ and development of a new medical technology which will be applicable for the cancer treatment.⁴ For such applications, especially, for spectroscopy, stabilization of frequency and amplitude are very important. We intend to make our gyrotrons more stable for such a purpose. In addition, applications of gyrotron to spectroscopy in many fields require a very narrow frequency bandwidth. In the past, frequency spectrum of gyrotrons developed for plasma diagnostics has already measured by a heterodyne detection system with a gunn diode as a reference oscillator.^{5,6} The measurement indicates the half value width Δf is several tens kHz. However, in this measurement, the bandwidth of a gunn diode is comparable with the observed Δf . Therefore, the real frequency width of the gyrotron cannot be seen. In order to develop a high stable gyrotron with a narrow frequency bandwidth, the accurate measurement technique of frequency is necessary.

In such a viewpoint, we tried to use a FIR molecular laser as a reference radiation source for the accurate measurement of frequency. Highly stable operation of FIR laser may be useful to measure the frequency of gyrotron output accurately. In this monograph, some preliminary experimen-

tal results for the accurate measurement of gyrotron frequency are described.

II. EXPERIMENTAL APPARATUS AND PROCEDURES

Figure 1 shows the experimental apparatus for accurate measurement of gyrotron frequency using a FIR laser as a reference. Gyrotron FU IV operates in cw mode. The operation frequency is around 295 GHz. The output power is around 20 W. It is transmitted through an oversized circular waveguide and fed on a Schottky barrier diode. The real power fed on the diode is 70 mW. The diode operates as a harmonic mixer. After mixing with the signal from a FIR

FIG. 1. Block diagram of frequency measurement system using a FIR laser as a reference.

FIG. 2. Frequency spectrum of Gyrotron FU IV output measured by using a synthesizer as a reference. Output frequency of the gyrotron $f_g = 295.2427$ GHz.

laser, an intermediate (IF) signal is observed on a spectrum analyzer. The IF signal includes a frequency component of four times of gyrotron frequency $4f_g$.

Here, we use a CD_3OH laser excited by a CO_2 laser. The frequency f_{FIR} is 1.1815889 THz. And the incident power fed onto the diode is 6 mW.

III. EXPERIMENTAL RESULTS

Before measurement of gyrotron frequency using a FIR laser, we tried it by the use of a synthesizer as a reference. Figure 2 shows a result of the measurement. This frequency spectrum is one of the IF signal, which is obtained by the mixing of gyrotron output with 17th harmonic of the synthesizer output. Therefore, the signal with the frequency $f = |f_g - 17f_{\text{thy}}|$ is analyzed. On the frequency spectrum, the half value width Δf can be observed. It is seen that Δf

around 20 kHz or larger. This frequency width of IF signal comes mainly from the frequency width of the synthesizer. In this measurement, both the feature of frequency width and frequency fluctuation of the gyrotron cannot be resolved so accurately.

Figure 3 shows frequency spectra of IF signal, which are observed by the use of a FIR laser as a reference. In Fig. 3(a), the total sweep time is 10 ms. In this case, the observed frequency width is smaller than 10 kHz. A real frequency width of gyrotron output Δf is one fourth of it. Therefore, Δf is around 2 kHz. This value is much smaller than the frequency width of resonant cavity. One possible explanation on the observed frequency width is the effect of technical noise suggested theoretically.⁷

In Fig. 3(b), the total sweep time is 10 s. In the figure, some frequency fluctuations are observed. The width of frequency fluctuation is around 40 kHz. So, the real frequency fluctuation of the gyrotron output is 10 kHz. Such a frequency fluctuation comes mainly from the fluctuation of high voltage power supplies for an anode and a cathode of electron gun. We should stabilize the power supplies for removing the frequency fluctuation. Highly stable operation of a FIR laser enables us to make an accurate measurement of gyrotron frequency. The frequency resolution of the measurement is much higher than 1 kHz. This is a convenient and useful tool for development of high quality gyrotrons with stable frequency.

IV. DISCUSSION

By the use of a FIR laser as a reference, we have succeeded the highly accurate measurement of gyrotron frequency. The frequency resolution of the measurement is higher than 1 kHz. Real frequency width Δf and the frequency fluctuation of the gyrotron are around 2 and 10 kHz, respectively. Such an accurate measurement of frequency will be a powerful tool for supporting our future development of high quality gyrotron with stable frequency, which is useful for spectroscopy in many fields.

FIG. 3. Frequency spectrum of gyrotron FU IV output measured by using a FIR laser as a reference: (a) full sweep time 10 ms and (b) 10 s.

ACKNOWLEDGMENT

The study was partially supported by Grant-in-Aid from Japan Society for Promotion of Science (JSPS).

¹T. Idehara, I. Ogawa, S. Mitsudo, M. Pereyaslavets, N. Nishida, and K. Yoshida, *IEEE Trans. Plasma Sci.* 27, 340 (1999).

²I. Ogawa, K. Yoshisue, H. Ibe, T. Idehara, and K. Kawahata, *Rev. Sci. Instrum.* 65, 1788 (1994).

³S. Mitsudo, Aripin, T. Shirai, T. Matsuda, T. Kanemaki, and T. Idehara, *Int. J. Infrared Millim. Waves* 21, 661 (2000).

⁴T. Tatsukawa, A. Doi, M. Teranaka, H. Takashima, F. Goda, T. Idehara, I. Ogawa, S. Mitsudo, and T. Kanemaki, *Int. J. Infrared Millim. Waves* 21, 1155 (2000).

⁵Y. Terumichi *et al.*, Conference Digest, 9th International Conference on Infrared and Millimeter Waves, Takarazuka, Japan, October 22–26, 1984, pp. 411–412.

⁶T. Idehara, T. Tatsukawa, H. Tanabe, S. Matsumoto, K. Kunieda, K. Hemmi, and T. Kanemaki, *Phys. Fluids B* 3, 1766 (1991).

⁷O. Dumbrajs and G. S. Nusinovich, *Phys. Plasmas* 4, 1413 (1997).